
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001759
7:04 15/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:57
STERILIZING 4:02
136.4 C 220kPa 4:02
136.2 C 222kPa 4:32
136.2 C 222kPa 5:02
136.2 C 222kPa 5:32
136.4 C 223kPa 6:02
136.4 C 223kPa 6:32
136.3 C 221kPa 7:02
136.4 C 222kPa 7:32
136.4 C 222kPa 7:32
VENTING 7:32
AIR DRYING 8:24
CYCLE COMPLETE 8:34

Digital Signature #
3715B50E82D847E5
