
Midmark M11 - v1.0.0
Total Cycles: 2854

Name: RITTER 1
Sterilizer ID: V2178198
02/03/2023 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.4 F 0.0
2:00 84.5 F 0.0
4:00 94.5 F 0.9
6:00 124.0 F 3.5
8:00 150.0 F 6.8
10:00 175.1 F 10.8
12:00 207.4 F 3.4
14:00 218.3 F 6.3
16:00 229.5 F 10.2
18:00 233.7 F 8.7
20:00 239.0 F 10.5
22:00 235.8 F 7.8
24:00 249.2 F 14.5
26:00 261.2 F 21.2
28:00 272.2 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.2 F 28.5
2:00 272.7 F 28.7
3:00 272.7 F 28.6
4:00 272.6 F 28.5

Min 271.6 F 27.9
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 236.7 F 0.0
Dry Cycle Aborted

FILLING: 0:56
HEATING: 28:21
STERILIZING: 4:00
VENTING: 2:49
DRYING: 2:58
TOTAL CYCLE: 00:39:04

02/03/2023 10:43 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____