

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001708  
16:22 21/02/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:32  
PRESSURIZING 2:38  
STERILIZING 3:44  
136.4 C 225kPa 3:44  
136.4 C 226kPa 4:14  
136.4 C 224kPa 4:44  
136.1 C 222kPa 5:14  
136.3 C 223kPa 5:44  
136.3 C 225kPa 6:14  
136.4 C 226kPa 6:44  
136.4 C 226kPa 7:14  
136.4 C 226kPa 7:14  
VENTING 7:14  
AIR DRYING 8:07  
CYCLE COMPLETE 8:13  
-----

Digital Signature #  
410AF3FC4AF2499E  
-----