

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4259

Name: RITTER2  
Sterilizer ID: V2313228  
01/13/2025 11:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.6 F 0.0  
2:00 70.5 F 0.0  
4:00 73.0 F 0.0  
6:00 93.4 F 0.0  
8:00 138.8 F 0.2  
10:00 164.7 F 0.5  
12:00 185.5 F 0.8  
14:00 199.1 F 3.3  
16:00 216.2 F 6.4  
18:00 221.2 F 9.9  
20:00 232.4 F 11.9  
22:00 234.9 F 9.7  
24:00 238.9 F 10.5  
26:00 244.0 F 12.9  
28:00 242.9 F 12.0  
30:00 254.6 F 18.1  
32:00 265.1 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.5 F 29.1  
2:00 272.5 F 29.6  
3:00 271.7 F 28.7  
4:00 272.6 F 29.3  
  
Min 271.5 F 28.7  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.9 F 0.0  
5:00 210.0 F 0.0  
10:00 187.1 F 0.0  
15:00 181.1 F 0.0  
20:00 191.8 F 0.0  
25:00 206.0 F 0.0  
30:00 211.0 F 0.0  
  
Drying Complete

FILLING: 0:50  
HEATING: 33:54  
STERILIZING: 4:00  
VENTING: 3:46  
DRYING: 30:00  
TOTAL CYCLE: 01:12:30  
CYCLE COMPLETE

01/13/2025 12:16 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_