
Midmark M11 - v1.0.0
Total Cycles: 644

Name: MIDMARK
Sterilizer ID: V2178198
02/01/2021 09:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.9 F 0.0
2:00 115.9 F 0.0
4:00 134.0 F 1.2
6:00 149.8 F 4.2
8:00 181.1 F 8.2
10:00 202.6 F 12.9
12:00 220.9 F 6.7
14:00 235.8 F 11.7
16:00 235.1 F 8.4
18:00 239.6 F 10.7
20:00 245.1 F 12.6
22:00 257.5 F 19.1
24:00 268.6 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 28.4
1:00 272.0 F 28.3
2:00 271.6 F 28.4
3:00 271.4 F 28.5
4:00 271.5 F 28.4

Min 271.4 F 28.3
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.7 F 0.0
5:00 238.2 F 0.0
10:00 229.8 F 0.0
15:00 228.0 F 0.0
20:00 225.0 F 0.0
25:00 223.9 F 0.0
30:00 199.7 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 24:55
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 01:02:46
CYCLE COMPLETE

02/01/2021 10:40 AM
ACCEPTED: X
REJECTED:
Operator: U01