
Midmark M11 - v1.0.5
Total Cycles: 1229

Name: RITTER2
Sterilizer ID: V2313228
06/27/2022 01:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.9 F 0.0
2:00 138.4 F 2.2
4:00 157.7 F 5.5
6:00 184.2 F 10.2
8:00 205.9 F 5.5
10:00 222.4 F 10.0
12:00 229.3 F 8.1
14:00 238.0 F 12.4
16:00 238.9 F 10.9
18:00 252.1 F 17.5
20:00 264.0 F 24.5
22:00 271.2 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.6
1:00 272.5 F 30.3
2:00 271.7 F 29.6
3:00 271.7 F 30.2
4:00 272.4 F 30.1

Min 271.2 F 29.4
Max 272.6 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 206.0 F 0.0
10:00 213.3 F 0.0
15:00 216.4 F 0.0
20:00 225.2 F 0.0
25:00 228.6 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 22:03
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:21
CYCLE COMPLETE

06/27/2022 02:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____