

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004993  
10:54 24/01/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:04  
PRESSURIZING 2:07  
STERILIZING 3:04  
136.4 C 223kPa 3:04  
135.9 C 220kPa 3:34  
136.2 C 223kPa 4:04  
136.4 C 224kPa 4:34  
136.3 C 222kPa 5:04  
136.4 C 223kPa 5:34  
136.4 C 224kPa 6:04  
136.4 C 221kPa 6:34  
136.4 C 221kPa 6:34  
VENTING 6:35  
AIR DRYING 7:47  
CYCLE COMPLETE 9:25

-----  
Digital Signature #  
4D56A2764EDF28F1  
-----