
Midmark M11 - v1.0.0
Total Cycles: 2013

Name: RITTER 1
Sterilizer ID: V2178198
06/09/2022 06:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.6 F 0.0
2:00 76.2 F 0.0
4:00 84.8 F 1.5
6:00 110.1 F 4.1
8:00 147.4 F 7.4
10:00 176.0 F 11.2
12:00 215.0 F 2.8
14:00 211.5 F 5.8
16:00 223.9 F 9.3
18:00 234.3 F 13.2
20:00 234.8 F 9.0
22:00 241.9 F 10.7
24:00 243.7 F 12.0
26:00 255.5 F 18.1
28:00 266.0 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 272.6 F 29.3
2:00 272.3 F 29.3
3:00 272.0 F 29.3
4:00 271.5 F 29.0

Min 270.6 F 28.2
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.5 F 0.0
5:00 218.7 F 0.0
10:00 171.0 F 0.0
15:00 184.0 F 0.0
20:00 187.3 F 0.0
25:00 191.5 F 0.0
30:00 183.1 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 29:36
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:07:37
CYCLE COMPLETE

06/09/2022 08:06 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____