
Midmark M11 - v1.0.0
Total Cycles: 2230

Name: RITTER 1
Sterilizer ID: V2178198
08/16/2022 12:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.5 F 0.0
2:00 199.7 F 4.6
4:00 206.0 F 10.0
6:00 222.7 F 5.6
8:00 234.9 F 10.4
10:00 234.1 F 7.7
12:00 245.4 F 13.4
14:00 244.2 F 12.3
16:00 257.4 F 19.2
18:00 268.9 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 28.7
2:00 271.8 F 28.7
3:00 272.5 F 28.6
4:00 271.8 F 28.7

Min 271.4 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.1 F 0.0
5:00 198.1 F 0.0
10:00 197.1 F 0.0
15:00 200.5 F 0.0
20:00 198.8 F 0.0
25:00 196.7 F 0.0
30:00 196.2 F 0.0
Drying Complete

FILLING: 2:25
HEATING: 19:05
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:43
CYCLE COMPLETE

08/16/2022 01:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____