
Midmark M11 - v1.0.0
Total Cycles: 5155

Name: RITTER 1
Sterilizer ID: V2178198
12/04/2024 11:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.4 F 0.0
2:00 169.0 F 0.2
4:00 180.4 F 0.8
6:00 212.1 F 2.2
8:00 226.7 F 6.5
10:00 232.0 F 11.0
12:00 230.3 F 7.0
14:00 244.4 F 13.0
16:00 242.6 F 11.5
18:00 240.5 F 10.4
20:00 254.4 F 17.3
22:00 265.9 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.4 F 28.8
2:00 272.3 F 28.5
3:00 271.7 F 28.0
4:00 271.7 F 28.3

Min 271.4 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.9 F 0.0
5:00 217.3 F 0.0
10:00 202.0 F 0.0
15:00 207.0 F 0.0
20:00 222.7 F 0.0
25:00 239.4 F 0.0
30:00 231.2 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 23:49
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:02:25
CYCLE COMPLETE

12/04/2024 12:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____