
Midmark M11 - v1.0.0
Total Cycles: 2947

Name: RITTER 1
Sterilizer ID: V2178198
02/27/2023 01:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.6 F 0.0
2:00 166.5 F 3.1
4:00 178.0 F 6.9
6:00 203.9 F 11.4
8:00 217.8 F 5.5
10:00 233.8 F 10.4
12:00 233.7 F 7.5
14:00 243.4 F 12.9
16:00 241.9 F 11.3
18:00 255.2 F 18.1
20:00 267.1 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 29.0
2:00 271.3 F 28.4
3:00 272.5 F 29.2
4:00 271.7 F 28.5

Min 271.3 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 224.2 F 0.0
10:00 225.9 F 0.0
15:00 222.7 F 0.0
20:00 227.0 F 0.0
25:00 228.1 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 21:26
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

02/27/2023 02:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____