
Midmark M11 - v1.0.0
Total Cycles: 1253

Name: RITTER 1
Sterilizer ID: V2178198
10/12/2021 08:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.8 F 0.0
2:00 106.0 F 0.0
4:00 117.3 F 0.1
6:00 138.6 F 0.7
8:00 165.4 F 3.1
10:00 189.1 F 5.9
12:00 203.4 F 9.1
14:00 216.7 F 12.8
16:00 225.3 F 7.3
18:00 236.1 F 11.9
20:00 234.9 F 9.0
22:00 240.4 F 10.5
24:00 245.7 F 13.1
26:00 257.0 F 19.1
28:00 267.3 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.4 F 29.0
2:00 271.7 F 28.3
3:00 271.6 F 28.7
4:00 272.3 F 28.9

Min 271.1 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 227.3 F 0.0
10:00 219.8 F 0.0
15:00 189.6 F 0.0
20:00 173.7 F 0.0
25:00 174.2 F 0.0
30:00 167.2 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:32
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:07:53
CYCLE COMPLETE

10/12/2021 10:02 AM
ACCEPTED: X
REJECTED:
Operator: U01