
Midmark M11 - v1.0.0
Total Cycles: 1084

Name: RITTER 1
Sterilizer ID: V2178198
08/10/2021 08:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.1 F 0.0
2:00 122.7 F 0.0
4:00 137.6 F 0.5
6:00 167.3 F 3.0
8:00 185.9 F 6.7
10:00 202.5 F 10.6
12:00 217.4 F 5.5
14:00 228.3 F 8.0
16:00 237.4 F 12.7
18:00 237.0 F 9.7
20:00 239.1 F 9.1
22:00 247.7 F 14.1
24:00 259.3 F 20.4
26:00 269.9 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.6 F 28.9
2:00 271.6 F 28.2
3:00 272.8 F 29.1
4:00 271.9 F 28.4

Min 271.6 F 28.2
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 218.1 F 0.0
10:00 217.3 F 0.0
15:00 198.0 F 0.0
20:00 193.7 F 0.0
25:00 198.6 F 0.0
30:00 205.5 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 26:57
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:05:06
CYCLE COMPLETE

08/10/2021 09:26 AM
ACCEPTED: X
REJECTED:
Operator: U01