
Midmark M11 - v1.0.0
Total Cycles: 2044

Name: RITTER 1
Sterilizer ID: V2178198
06/22/2022 10:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.2 F 0.0
2:00 188.3 F 3.5
4:00 192.0 F 7.4
6:00 215.2 F 12.0
8:00 225.4 F 6.5
10:00 235.8 F 11.8
12:00 234.9 F 8.8
14:00 240.5 F 10.1
16:00 246.3 F 13.4
18:00 258.1 F 19.9
20:00 269.0 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.6 F 29.1
2:00 271.5 F 28.3
3:00 272.6 F 29.2
4:00 272.1 F 28.8

Min 271.4 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 231.6 F 0.0
10:00 226.4 F 0.0
15:00 231.1 F 0.0
20:00 232.9 F 0.0
25:00 225.8 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:08
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:24
CYCLE COMPLETE

06/22/2022 11:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____