
Midmark M11 - v1.0.5
Total Cycles: 3328

Name: RITTER2
Sterilizer ID: V2313228
01/26/2024 04:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.6 F 0.0
2:00 145.1 F 2.0
4:00 163.8 F 5.4
6:00 191.4 F 9.8
8:00 213.1 F 5.9
10:00 223.7 F 8.7
12:00 233.4 F 11.9
14:00 238.7 F 11.1
16:00 237.2 F 9.6
18:00 251.3 F 16.4
20:00 263.2 F 23.4
22:00 271.8 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.4 F 29.4
2:00 271.4 F 29.1
3:00 272.5 F 29.5
4:00 271.3 F 28.9

Min 271.3 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.9 F 0.0
5:00 237.7 F 0.0
10:00 217.1 F 0.0
15:00 203.9 F 0.0
20:00 196.4 F 0.0
25:00 197.2 F 0.0
30:00 197.2 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 22:07
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:00:35
CYCLE COMPLETE

01/26/2024 05:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____