
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001423
12:28 14/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:23
PRESSURIZING 3:25
STERILIZING 4:31
136.0 C 222kPa 4:31
136.0 C 222kPa 5:01
136.3 C 223kPa 5:31
136.3 C 224kPa 6:01
136.4 C 225kPa 6:31
136.3 C 223kPa 7:01
136.1 C 223kPa 7:31
136.3 C 225kPa 8:01
136.3 C 225kPa 8:01
VENTING 8:01
AIR DRYING 8:54
CYCLE COMPLETE 20:11

Digital Signature #
6D63D9D18600BB0F
