
Midmark M11 - v1.0.5
Total Cycles: 4468

Name: RITTER2
Sterilizer ID: V2313228
03/27/2025 05:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.2 F 0.0
2:00 175.0 F 3.2
4:00 192.0 F 8.0
6:00 209.5 F 12.8
8:00 220.0 F 8.4
10:00 232.1 F 12.8
12:00 237.8 F 11.4
14:00 238.5 F 10.2
16:00 252.1 F 17.0
18:00 263.4 F 23.7
20:00 272.5 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.2 F 30.4
2:00 272.3 F 29.9
3:00 272.8 F 30.4
4:00 272.0 F 30.3

Min 270.9 F 29.1
Max 273.4 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.8 F 0.0
5:00 204.1 F 0.0
10:00 203.8 F 0.0
15:00 202.9 F 0.0
20:00 200.1 F 0.0
25:00 198.2 F 0.0
30:00 196.9 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 20:08
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 00:59:55
CYCLE COMPLETE

03/27/2025 06:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____