

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 001539  
9:48 24/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:20  
PRESSURIZING 3:20  
STERILIZING 4:30  
136.4 C 224kPa 4:30  
136.3 C 225kPa 5:00  
136.4 C 224kPa 5:30  
136.0 C 222kPa 6:00  
136.3 C 223kPa 6:30  
136.2 C 223kPa 7:00  
136.4 C 225kPa 7:30  
136.3 C 225kPa 8:00  
136.3 C 225kPa 8:00  
VENTING 8:00  
AIR DRYING 8:51  
CYCLE COMPLETE 11:26  
-----

Digital Signature #  
5CC400032A06466A  
-----