

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001188  
10:53 22/07/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:35  
PRESSURIZING 2:42  
STERILIZING 3:52  
136.0 C 220kPa 3:52  
135.9 C 223kPa 4:22  
136.3 C 225kPa 4:52  
136.3 C 225kPa 5:22  
136.3 C 225kPa 5:52  
136.2 C 223kPa 6:22  
136.4 C 223kPa 6:52  
136.2 C 221kPa 7:22  
136.2 C 221kPa 7:22  
VENTING 7:22  
AIR DRYING 8:15  
CYCLE COMPLETE 23:15  
-----

Digital Signature #  
240D6514DF245026  
-----