
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000424
13:37 08/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:40
PRESSURIZING 3:41
STERILIZING 4:43
136.4 C 224kPa 4:43
136.1 C 223kPa 5:13
136.4 C 225kPa 5:43
136.5 C 225kPa 6:13
136.4 C 222kPa 6:43
136.3 C 223kPa 7:13
136.1 C 222kPa 7:43
136.4 C 223kPa 8:13
136.4 C 223kPa 8:13
VENTING 8:13
AIR DRYING 9:06
CYCLE COMPLETE 9:18

Digital Signature #
DDDF87CF3A17C224
