
Midmark M11 - v1.0.0
Total Cycles: 3888

Name: RITTER 1
Sterilizer ID: V2178198
10/11/2023 10:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.2 F 0.0
2:00 189.2 F 3.6
4:00 192.3 F 8.2
6:00 223.3 F 13.1
8:00 227.2 F 7.5
10:00 239.5 F 12.7
12:00 237.8 F 9.9
14:00 237.4 F 8.5
16:00 249.6 F 15.0
18:00 261.9 F 21.9
20:00 272.8 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.2
1:00 272.4 F 28.9
2:00 271.2 F 28.1
3:00 272.4 F 29.0
4:00 271.7 F 28.3

Min 271.0 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 230.0 F 0.0
10:00 246.1 F 0.0
15:00 238.7 F 0.0
20:00 230.0 F 0.0
25:00 223.9 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 20:21
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:59:04
CYCLE COMPLETE

10/11/2023 11:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____