
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000572
12:18 09/08/2021
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:09
PRESSURIZING 4:54
STERILIZING 6:09
136.4 C 232kPa 6:09
136.4 C 232kPa 6:39
136.4 C 230kPa 7:09
136.2 C 230kPa 7:39
136.3 C 232kPa 8:09
136.3 C 230kPa 8:39
136.2 C 230kPa 9:09
136.4 C 232kPa 9:39
136.4 C 233kPa 10:09
136.1 C 230kPa 10:39
136.4 C 232kPa 11:09
136.1 C 229kPa 11:39
136.3 C 231kPa 12:09
136.4 C 233kPa 12:39
136.1 C 229kPa 13:09
136.2 C 231kPa 13:39
136.2 C 231kPa 14:09
136.4 C 233kPa 14:39
136.4 C 233kPa 15:09
136.4 C 232kPa 15:39
136.4 C 233kPa 16:09
136.4 C 233kPa 16:09
VENTING 16:09
AIR DRYING 17:03
CYCLE COMPLETE 32:03

Digital Signature #
3A8A805FE47B4AC0
