
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000728
10:54 22/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:11
PRESSURIZING 2:15
STERILIZING 3:17
136.0 C 228kPa 3:17
136.4 C 233kPa 3:47
136.2 C 231kPa 4:17
136.4 C 233kPa 4:47
136.3 C 233kPa 5:17
136.4 C 233kPa 5:47
136.4 C 233kPa 6:17
136.3 C 232kPa 6:47
136.3 C 232kPa 6:47
VENTING 6:47
AIR DRYING 7:42
CYCLE COMPLETE 7:45

Digital Signature #
0A744E4935DBBBEE
