
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000614
15:00 27/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:20
STERILIZING 4:27
136.0 C 229kPa 4:27
136.3 C 232kPa 4:57
136.3 C 233kPa 5:27
136.2 C 229kPa 5:57
136.3 C 233kPa 6:27
136.3 C 233kPa 6:57
136.3 C 230kPa 7:27
136.3 C 232kPa 7:57
136.3 C 232kPa 7:57
VENTING 7:57
AIR DRYING 8:50
CYCLE COMPLETE 22:20

Digital Signature #
03E9C2A00901F1E2
