
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000759
14:43 02/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:35
PRESSURIZING 3:36
STERILIZING 4:50
136.4 C 223kPa 4:50
136.3 C 225kPa 5:20
136.2 C 223kPa 5:50
136.3 C 223kPa 6:20
136.0 C 223kPa 6:50
136.4 C 225kPa 7:20
136.4 C 225kPa 7:50
136.4 C 225kPa 8:20
136.4 C 225kPa 8:20
VENTING 8:20
AIR DRYING 9:14
CYCLE COMPLETE 9:17

Digital Signature #
5DAC15DEB2245733
