
Midmark M11 - v1.0.0
Total Cycles: 5363

Name: RITTER 1
Sterilizer ID: V2178198
02/25/2025 12:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 100.6 F 0.0
2:00 106.8 F 0.9
4:00 116.1 F 2.9
6:00 142.8 F 5.7
8:00 168.2 F 9.0
10:00 190.7 F 12.6
12:00 207.1 F 3.7
14:00 214.9 F 7.0
16:00 227.8 F 10.8
18:00 231.7 F 7.4
20:00 239.1 F 11.5
22:00 236.3 F 8.9
24:00 248.5 F 14.6
26:00 259.0 F 20.3
28:00 268.5 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.2 F 29.4
2:00 272.2 F 29.1
3:00 271.3 F 28.6
4:00 271.3 F 28.9

Min 271.0 F 28.5
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.4 F 0.0
5:00 198.2 F 0.0
10:00 204.9 F 0.0
15:00 200.9 F 0.0
20:00 201.3 F 0.0
25:00 207.6 F 0.0
30:00 210.3 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 29:22
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:08:05
CYCLE COMPLETE

02/25/2025 01:34 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____