

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2359

Name: RITTER2  
Sterilizer ID: V2313228  
05/12/2023 09:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.1 F 0.0  
2:00 184.3 F 2.1  
4:00 181.9 F 5.5  
6:00 201.3 F 10.4  
8:00 217.7 F 5.6  
10:00 229.6 F 10.5  
12:00 231.6 F 7.9  
14:00 244.0 F 13.5  
16:00 243.2 F 12.4  
18:00 255.6 F 18.8  
20:00 266.6 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.9  
1:00 272.1 F 29.8  
2:00 272.4 F 29.6  
3:00 271.9 F 29.2  
4:00 271.1 F 28.8  
  
Min 271.1 F 28.8  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.0 F 0.0  
5:00 209.9 F 0.0  
10:00 217.1 F 0.0  
15:00 220.1 F 0.0  
20:00 218.7 F 0.0  
25:00 219.3 F 0.0  
30:00 218.5 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 21:43  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:00:36  
CYCLE COMPLETE

05/12/2023 10:12 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_