
Midmark M11 - v1.0.5
Total Cycles: 1033

Name: RITTER2
Sterilizer ID: V2313228
04/26/2022 09:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.9 F 0.0
2:00 165.0 F 3.0
4:00 181.2 F 6.9
6:00 203.5 F 12.0
8:00 221.4 F 6.6
10:00 236.7 F 12.0
12:00 235.7 F 9.4
14:00 238.1 F 9.4
16:00 248.4 F 14.9
18:00 260.6 F 21.7
20:00 271.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.6 F 29.7
2:00 271.6 F 28.9
3:00 272.0 F 29.8
4:00 272.7 F 29.4

Min 271.1 F 28.9
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.4 F 0.0
5:00 233.9 F 0.0
10:00 206.3 F 0.0
15:00 208.1 F 0.0
20:00 189.2 F 0.0
25:00 189.5 F 0.0
30:00 173.3 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 20:38
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:04
CYCLE COMPLETE

04/26/2022 10:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____