
Midmark M11 - v1.0.0
Total Cycles: 4681

Name: RITTER 1
Sterilizer ID: V2178198
06/26/2024 03:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.8 F 0.0
2:00 153.8 F 2.7
4:00 168.1 F 6.2
6:00 199.1 F 10.7
8:00 214.8 F 4.8
10:00 229.0 F 10.1
12:00 230.7 F 7.7
14:00 240.8 F 12.7
16:00 240.2 F 11.2
18:00 253.5 F 17.8
20:00 265.2 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.6 F 29.5
2:00 271.2 F 29.1
3:00 272.4 F 29.5
4:00 271.2 F 28.9

Min 271.1 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 223.3 F 0.0
10:00 235.1 F 0.0
15:00 229.8 F 0.0
20:00 229.2 F 0.0
25:00 228.7 F 0.0
30:00 221.0 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:00:34
CYCLE COMPLETE

06/26/2024 04:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____