
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006058
8:00 02/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:05
STERILIZING 4:12
136.4 C 222kPa 4:12
136.3 C 223kPa 4:42
136.4 C 219kPa 5:12
136.0 C 221kPa 5:42
136.2 C 220kPa 6:12
136.4 C 221kPa 6:42
136.3 C 221kPa 7:12
136.4 C 222kPa 7:42
136.4 C 222kPa 7:42
VENTING 7:42
AIR DRYING 8:31
CYCLE COMPLETE 10:00

Digital Signature #
48DF0D63A8BD83F7
