

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 008419  
7:37 19/03/2025

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:10  
PRESSURIZING 3:12  
STERILIZING 4:09  
136.2 C 220kPa 4:09  
136.1 C 220kPa 4:39  
136.4 C 220kPa 5:09  
136.2 C 219kPa 5:39  
136.4 C 219kPa 6:09  
136.4 C 220kPa 6:39  
136.2 C 218kPa 7:09  
136.3 C 218kPa 7:39  
136.3 C 218kPa 7:39  
VENTING 7:39  
AIR DRYING 8:37  
CYCLE COMPLETE 8:49

-----  
Digital Signature #  
66DD0077AA8F5026  
-----