

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1199

Name: RITTER 1  
Sterilizer ID: V2178198  
09/23/2021 09:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.8 F 0.0  
2:00 175.7 F 2.9  
4:00 185.0 F 6.4  
6:00 204.9 F 10.9  
8:00 223.0 F 4.9  
10:00 235.9 F 10.7  
12:00 233.1 F 7.5  
14:00 243.7 F 12.0  
16:00 245.0 F 12.6  
18:00 257.4 F 19.3  
20:00 268.6 F 26.2  
22:00 271.8 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.7  
1:00 272.9 F 29.2  
2:00 271.7 F 28.8  
3:00 272.4 F 29.2  
4:00 271.6 F 28.7

Min 271.6 F 28.6  
Max 273.1 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.5 F 0.0  
5:00 236.4 F 0.0  
10:00 221.7 F 0.0  
15:00 221.2 F 0.0  
20:00 216.6 F 0.0  
25:00 216.6 F 0.0  
30:00 212.4 F 0.0  
Drying Complete

FILLING: 1:44  
HEATING: 22:03  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:00:55  
CYCLE COMPLETE

09/23/2021 10:49 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_