
Midmark M11 - v1.0.0
Total Cycles: 3720

Name: RITTER 1
Sterilizer ID: V2178198
09/05/2023 03:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.7 F 0.0
2:00 195.1 F 3.6
4:00 217.1 F 9.4
6:00 224.7 F 6.4
8:00 237.6 F 11.2
10:00 236.2 F 9.0
12:00 238.9 F 9.3
14:00 249.9 F 15.2
16:00 262.6 F 22.4
18:00 272.1 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.3 F 28.6
2:00 272.3 F 28.9
3:00 271.9 F 28.3
4:00 272.6 F 29.1

Min 271.5 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 236.6 F 0.0
10:00 224.3 F 0.0
15:00 232.6 F 0.0
20:00 226.8 F 0.0
25:00 226.7 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 18:08
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:34
CYCLE COMPLETE

09/05/2023 03:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____