
Midmark M11 - v1.0.5
Total Cycles: 429

Name: RITTER2
Sterilizer ID: V2313228
10/07/2021 03:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.6 F 0.0
2:00 172.7 F 3.9
4:00 196.6 F 9.0
6:00 218.3 F 7.4
8:00 228.9 F 10.3
10:00 232.9 F 8.1
12:00 242.1 F 11.8
14:00 245.4 F 13.4
16:00 258.3 F 20.2
18:00 269.5 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 271.7 F 29.2
2:00 272.1 F 29.6
3:00 272.4 F 29.6
4:00 272.3 F 29.5

Min 271.3 F 28.8
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.7 F 0.0
5:00 230.7 F 0.0
10:00 225.4 F 0.0
15:00 219.4 F 0.0
20:00 209.2 F 0.0
25:00 205.2 F 0.0
30:00 204.1 F 0.0
Drying Complete

FILLING: 2:11
HEATING: 18:53
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:14
CYCLE COMPLETE

10/07/2021 04:36 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____