
Midmark M11 - v1.0.5
Total Cycles: 2302

Name: RITTER2
Sterilizer ID: V2313228
04/25/2023 02:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.1 F 0.0
2:00 177.8 F 4.1
4:00 193.2 F 9.0
6:00 213.3 F 7.7
8:00 227.0 F 9.6
10:00 233.4 F 8.6
12:00 242.4 F 12.0
14:00 246.1 F 13.8
16:00 259.5 F 21.1
18:00 271.5 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.7 F 29.6
2:00 271.5 F 28.8
3:00 272.5 F 29.8
4:00 272.4 F 29.3

Min 271.3 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.3 F 0.0
5:00 209.5 F 0.0
10:00 216.5 F 0.0
15:00 211.0 F 0.0
20:00 232.5 F 0.0
25:00 226.3 F 0.0
30:00 217.1 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 18:38
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:42
CYCLE COMPLETE

04/25/2023 03:54 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____