

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5122

Name: RITTER 1  
Sterilizer ID: V2178198  
11/21/2024 03:26 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.6 F 0.0  
2:00 154.8 F 0.1  
4:00 178.2 F 0.6  
6:00 202.6 F 3.0  
8:00 222.6 F 7.0  
10:00 237.3 F 11.1  
12:00 228.2 F 6.8  
14:00 242.6 F 12.5  
16:00 239.8 F 10.7  
18:00 237.5 F 9.4  
20:00 251.6 F 16.2  
22:00 263.1 F 22.9  
24:00 272.3 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.9 F 29.2  
2:00 272.3 F 28.7  
3:00 271.5 F 28.4  
4:00 272.7 F 29.2

Min 271.5 F 28.3  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.9 F 0.0  
5:00 232.2 F 0.0  
10:00 231.2 F 0.0  
15:00 231.6 F 0.0  
20:00 226.8 F 0.0  
25:00 236.7 F 0.0  
30:00 224.5 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 24:14  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:02:56  
CYCLE COMPLETE

11/21/2024 04:28 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_