
Midmark M11 - v1.0.0
Total Cycles: 4787

Name: RITTER 1
Sterilizer ID: V2178198
08/07/2024 09:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.4 F 0.0
2:00 153.3 F 1.4
4:00 165.1 F 4.4
6:00 211.2 F 7.8
8:00 202.9 F 12.0
10:00 220.1 F 4.9
12:00 229.4 F 9.1
14:00 241.1 F 13.4
16:00 238.4 F 10.0
18:00 237.2 F 8.9
20:00 247.0 F 13.5
22:00 257.7 F 19.3
24:00 267.7 F 25.5
26:00 272.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.7
1:00 272.4 F 28.7
2:00 272.4 F 29.0
3:00 272.3 F 29.2
4:00 271.8 F 28.8

Min 271.7 F 28.6
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.6 F 0.0
5:00 211.9 F 0.0
10:00 203.9 F 0.0
15:00 208.3 F 0.0
20:00 203.3 F 0.0
25:00 201.7 F 0.0
30:00 200.0 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 26:09
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:04:39
CYCLE COMPLETE

08/07/2024 10:12 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____