
Midmark M11 - v1.0.0
Total Cycles: 5232

Name: RITTER 1
Sterilizer ID: V2178198
01/10/2025 09:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 72.1 F 0.0
4:00 86.2 F 0.0
6:00 113.6 F 0.1
8:00 149.5 F 0.2
10:00 174.7 F 0.4
12:00 198.2 F 0.9
14:00 217.4 F 4.6
16:00 223.3 F 7.8
18:00 232.8 F 11.3
20:00 229.3 F 6.5
22:00 242.7 F 12.2
24:00 239.5 F 10.0
26:00 236.6 F 8.3
28:00 250.0 F 14.9
30:00 261.7 F 21.5
32:00 272.1 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.9
1:00 271.9 F 27.9
2:00 272.3 F 28.2
3:00 272.7 F 28.5
4:00 272.2 F 28.4

Min 271.5 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.2 F 0.0
5:00 216.8 F 0.0
10:00 214.8 F 0.0
15:00 229.5 F 0.0
20:00 214.4 F 0.0
25:00 211.2 F 0.0
30:00 209.3 F 0.0

Drying Complete

FILLING: 0:42
HEATING: 32:26
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:10:38
CYCLE COMPLETE

01/10/2025 10:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____