

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3843

Name: RITTER2  
Sterilizer ID: V2313228  
08/08/2024 03:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.6 F 0.0  
2:00 170.0 F 2.7  
4:00 193.5 F 7.0  
6:00 212.9 F 11.3  
8:00 221.0 F 6.1  
10:00 232.1 F 10.8  
12:00 232.2 F 7.7  
14:00 243.5 F 13.4  
16:00 243.2 F 12.3  
18:00 254.7 F 18.3  
20:00 265.0 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.3 F 29.4  
2:00 271.8 F 29.7  
3:00 271.7 F 29.1  
4:00 272.1 F 29.8

Min 271.1 F 29.0  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.4 F 0.0  
5:00 214.0 F 0.0  
10:00 216.0 F 0.0  
15:00 216.0 F 0.0  
20:00 212.2 F 0.0  
25:00 212.3 F 0.0  
30:00 208.9 F 0.0  
Drying Complete

FILLING: 1:54  
HEATING: 21:54  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:01:34  
CYCLE COMPLETE

08/08/2024 04:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_