
Midmark M11 - v1.0.5
Total Cycles: 1959

Name: RITTER2
Sterilizer ID: V2313228
01/23/2023 01:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.8 F 0.0
2:00 154.8 F 0.1
4:00 169.0 F 0.8
6:00 195.7 F 4.9
8:00 212.5 F 9.3
10:00 222.7 F 9.0
12:00 233.5 F 10.8
14:00 233.5 F 9.1
16:00 237.4 F 9.7
18:00 250.3 F 15.9
20:00 263.0 F 23.3
22:00 271.8 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.8 F 29.5
2:00 271.9 F 28.8
3:00 272.2 F 29.5
4:00 272.6 F 29.4

Min 271.5 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.2 F 0.0
5:00 221.4 F 0.0
10:00 235.1 F 0.0
15:00 229.4 F 0.0
20:00 228.4 F 0.0
25:00 228.6 F 0.0
30:00 231.7 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 22:06
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:00:29
CYCLE COMPLETE

01/23/2023 02:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____