

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.6uS / 0.3ppm  
CYCLE NUMBER 000819  
9:31 13/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:47  
PRESSURIZING 3:48  
STERILIZING 4:54  
136.4 C 224kPa 4:54  
136.3 C 224kPa 5:24  
136.4 C 225kPa 5:54  
136.3 C 223kPa 6:24  
136.3 C 223kPa 6:54  
136.3 C 223kPa 7:24  
136.0 C 221kPa 7:54  
136.4 C 222kPa 8:24  
136.4 C 222kPa 8:24  
VENTING 8:24  
AIR DRYING 9:17  
CYCLE COMPLETE 10:03

-----  
Digital Signature #  
02F331E41BBF3B8F  
-----