

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3589

Name: RITTER 1  
Sterilizer ID: V2178198  
07/27/2023 09:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.0 F 0.0  
2:00 184.7 F 1.5  
4:00 195.0 F 5.0  
6:00 224.2 F 10.1  
8:00 227.8 F 6.0  
10:00 235.4 F 10.5  
12:00 236.2 F 7.5  
14:00 244.7 F 13.3  
16:00 243.0 F 11.9  
18:00 255.8 F 18.5  
20:00 267.1 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.1 F 28.7  
2:00 272.5 F 29.2  
3:00 271.5 F 28.3  
4:00 272.5 F 29.0  
  
Min 271.5 F 28.3  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.1 F 0.0  
5:00 227.4 F 0.0  
10:00 236.0 F 0.0  
15:00 230.0 F 0.0  
20:00 226.2 F 0.0  
25:00 225.4 F 0.0  
30:00 224.9 F 0.0  
Drying Complete

FILLING: 1:34  
HEATING: 21:25  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:00:10  
CYCLE COMPLETE

07/27/2023 10:56 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_