
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001426
8:59 14/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:52
PRESSURIZING 3:02
STERILIZING 4:13
136.0 C 220kPa 4:13
136.2 C 222kPa 4:43
136.1 C 222kPa 5:13
136.4 C 223kPa 5:43
136.4 C 224kPa 6:13
136.4 C 224kPa 6:43
136.3 C 223kPa 7:13
136.4 C 225kPa 7:43
136.4 C 225kPa 7:43
VENTING 7:43
AIR DRYING 8:31
CYCLE COMPLETE 9:10

Digital Signature #
4D722B20700A7B3B
