
Midmark M11 - v1.0.0
Total Cycles: 4187

Name: RITTER 1
Sterilizer ID: V2178198
01/10/2024 10:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.9 F 0.0
2:00 197.8 F 3.0
4:00 206.3 F 7.5
6:00 214.0 F 13.0
8:00 226.0 F 7.8
10:00 236.2 F 13.0
12:00 237.9 F 10.5
14:00 236.6 F 8.8
16:00 249.7 F 15.5
18:00 261.4 F 22.1
20:00 272.1 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.5 F 29.3
2:00 271.5 F 28.5
3:00 272.3 F 29.4
4:00 272.1 F 29.0

Min 271.2 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 232.1 F 0.0
10:00 228.0 F 0.0
15:00 229.3 F 0.0
20:00 230.0 F 0.0
25:00 226.3 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 00:59:55
CYCLE COMPLETE

01/10/2024 11:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____