

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2325

Name: RITTER 1  
Sterilizer ID: V2178198  
09/14/2022 12:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 133.8 F 0.0  
2:00 143.6 F 2.3  
4:00 157.0 F 5.6  
6:00 181.7 F 9.7  
8:00 203.6 F 8.4  
10:00 218.7 F 6.9  
12:00 229.2 F 11.4  
14:00 230.0 F 7.3  
16:00 242.2 F 13.0  
18:00 240.7 F 11.0  
20:00 253.2 F 17.4  
22:00 264.6 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.7 F 29.5  
2:00 271.7 F 28.7  
3:00 272.4 F 29.5  
4:00 272.4 F 29.2

Min 271.4 F 28.6  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.5 F 0.0  
5:00 215.1 F 0.0  
10:00 223.9 F 0.0  
15:00 238.7 F 0.0  
20:00 227.0 F 0.0  
25:00 228.4 F 0.0  
30:00 225.6 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 23:58  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:02:02  
CYCLE COMPLETE

09/14/2022 01:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_