
Midmark M11 - v1.0.5
Total Cycles: 585

Name: RITTER2
Sterilizer ID: V2313228
11/30/2021 07:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.6 F 0.0
2:00 72.9 F 0.0
4:00 78.0 F 0.0
6:00 112.3 F 0.1
8:00 143.1 F 0.2
10:00 173.4 F 0.5
12:00 198.5 F 3.3
14:00 212.7 F 6.9
16:00 223.2 F 10.5
18:00 228.8 F 8.6
20:00 236.6 F 10.3
22:00 235.7 F 8.9
24:00 244.6 F 13.2
26:00 243.1 F 12.0
28:00 254.9 F 18.2
30:00 265.5 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.2 F 29.0
2:00 272.5 F 29.5
3:00 271.6 F 28.7
4:00 272.3 F 29.2

Min 271.5 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.5 F 0.0
5:00 231.2 F 0.0
10:00 199.2 F 0.0
15:00 211.3 F 0.0
20:00 207.5 F 0.0
25:00 204.0 F 0.0
30:00 190.6 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 31:46
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:09:58
CYCLE COMPLETE

11/30/2021 09:39 AM
ACCEPTED: X
REJECTED:
Operator: U02