
Midmark M11 - v1.0.5
Total Cycles: 3616

Name: RITTER2
Sterilizer ID: V2313228
05/14/2024 02:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.7 F 0.0
2:00 99.1 F 0.2
4:00 107.5 F 1.7
6:00 141.9 F 4.3
8:00 168.5 F 7.3
10:00 186.6 F 10.6
12:00 211.6 F 7.2
14:00 217.1 F 7.1
16:00 226.1 F 10.8
18:00 228.6 F 8.2
20:00 238.3 F 11.6
22:00 236.8 F 9.6
24:00 248.7 F 15.3
26:00 259.3 F 21.1
28:00 269.1 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.7
1:00 272.2 F 29.7
2:00 272.3 F 30.2
3:00 271.5 F 29.9
4:00 272.2 F 29.8

Min 271.3 F 29.5
Max 272.9 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.0 F 0.0
5:00 228.1 F 0.0
10:00 189.0 F 0.0
15:00 185.7 F 0.0
20:00 190.7 F 0.0
25:00 194.2 F 0.0
30:00 189.3 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 29:58
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:08:40
CYCLE COMPLETE

05/14/2024 03:38 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____