
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002322
13:44 30/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:47
PRESSURIZING 1:51
STERILIZING 2:49
136.4 C 221kPa 2:49
136.0 C 222kPa 3:19
136.3 C 224kPa 3:49
136.4 C 222kPa 4:19
136.2 C 220kPa 4:49
136.3 C 224kPa 5:19
136.0 C 221kPa 5:49
136.3 C 223kPa 6:19
136.3 C 223kPa 6:19
VENTING 6:19
AIR DRYING 7:15
CYCLE COMPLETE 7:26

Digital Signature #
19E16671A693D0AC
