
Midmark M11 - v1.0.5
Total Cycles: 3044

Name: RITTER2
Sterilizer ID: V2313228
10/23/2023 02:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.4 F 0.0
2:00 124.9 F 1.0
4:00 145.9 F 4.0
6:00 172.7 F 7.7
8:00 193.4 F 11.9
10:00 211.8 F 4.7
12:00 223.7 F 9.1
14:00 232.8 F 11.7
16:00 237.3 F 10.8
18:00 235.8 F 9.0
20:00 249.3 F 15.5
22:00 260.7 F 22.0
24:00 271.3 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.7 F 29.9
2:00 272.1 F 29.3
3:00 271.5 F 29.3
4:00 272.5 F 30.0

Min 271.4 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 227.4 F 0.0
10:00 229.0 F 0.0
15:00 220.3 F 0.0
20:00 217.5 F 0.0
25:00 218.7 F 0.0
30:00 217.7 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 24:43
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:03:12
CYCLE COMPLETE

10/23/2023 03:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____