

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1949

Name: RITTER 1  
Sterilizer ID: V2178198  
05/25/2022 08:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 136.3 F 0.0  
2:00 138.6 F 1.7  
4:00 149.1 F 4.8  
6:00 174.4 F 8.5  
8:00 210.7 F 13.3  
10:00 216.5 F 7.3  
12:00 231.1 F 12.2  
14:00 231.3 F 8.1  
16:00 243.4 F 12.6  
18:00 242.1 F 11.5  
20:00 254.6 F 17.9  
22:00 266.1 F 24.5  
24:00 272.0 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.9  
1:00 272.8 F 29.3  
2:00 272.1 F 29.4  
3:00 272.4 F 29.0  
4:00 272.5 F 29.6

Min 271.8 F 28.8  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.3 F 0.0  
5:00 196.3 F 0.0  
10:00 220.4 F 0.0  
15:00 234.3 F 0.0  
20:00 233.7 F 0.0  
25:00 229.4 F 0.0  
30:00 228.4 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 24:30  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:02:28  
CYCLE COMPLETE

05/25/2022 09:58 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_